

## DECLARATION OF CONFORMITY

Manufacturer: CRYOLOGIC PTY. LTD.  
Address: 54 Geddes St, Mulgrave,  
Victoria 3170, AUSTRALIA.

Declares that the product:

Type: FREEZE CONTROL  
Model: CL-8800i SYSTEM  
Description: System for controlled rate freezing of biological specimens

Conforms to the following European Council Directive(s):

93 / 42 / EEC (Medical Device Directive - Class I)  
89 / 336 / EEC (Electromagnetic Compatibility)  
73 / 23 / EEC (Low Voltage)

Standards applied: EN 60601-1  
EN 55011  
EN 50082-1  
EN 61000-3-2  
EN 61000-3-3

Authorised by: Willem Lindemans  
Research Director,  
Cryologic Pty Ltd  
On behalf of Cryologic

